



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:*

### ***Ipsen USA***

**984 Ipsen Road, Cherry Valley, IL 61016**

*and hereby declares that the Organization is accredited in accordance with  
the recognized International Standard:*

### **ISO/IEC 17025:2017**

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

### ***Electrical Calibration (As detailed in the supplement)***

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

*Initial Accreditation Date:*

*Issue Date:*

*Expiration Date:*

Septemebr 26, 2023

November 12, 2025

December 31, 2027

*Accreditation No.:*

*Certificate No.:*

118324

L25-836

Tracy Szerszen  
President

*The validity of this certificate is maintained through ongoing assessments based  
on a continuous accreditation cycle. The validity of this certificate should be  
confirmed through the PJLA website: [www.pjlabs.com](http://www.pjlabs.com)*

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084




# Certificate of Accreditation: Supplement

## Ipsen USA

984 Ipsen Road, Cherry Valley, IL 61016  
Contact Name: Cavan Cardenas Phone: 815-979-9101

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ( $\pm$ ) <sup>1</sup>	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Source Thermocouple Type K	200 °F to 2 400 °F	0.48 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Source Thermocouple Type N	200 °F to 2 200 °F	0.37 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Source Thermocouple Type S	800 °F to 2 400 °F	1.5 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Read Thermocouple Type K	200 °F to 2 400 °F	0.44 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Read Thermocouple Type N	200 °F to 2 200 °F	0.25 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Multifunction Calibrator – Read Thermocouple Type S	800 °F to 2 400 °F	0.72 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F



# Certificate of Accreditation: Supplement

## Ipsen USA

984 Ipsen Road, Cherry Valley, IL 61016  
Contact Name: Cavan Cardenas Phone: 815-979-9101

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ( $\pm$ ) <sup>1</sup>	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Temperature Calibration, Indication, and Control Equipment used with Data Logger – Read Thermocouple Type K	200 °F to 2 400 °F	0.67 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F
Electrical	Temperature Calibration, Indication, and Control Equipment used with Data Logger – Read Thermocouple Type N	200 °F to 2 200 °F	0.54 °F	Electrical Simulation of Thermocouple Fluke 7526A	AMS 2750 IWI-19-01 AMS 2750	F1, F2, F3	F

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
- The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
- Location of activity:  

<b>Location Code</b>	<b>Location</b>
F	Conformity assessment activity is performed at the CABs fixed facility
- Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.